

Hellenic Accreditation System



Annex F2/21 to the Certificate No. **90-9**

SCOPE of ACCREDITATION

of the

Calibration Laboratory

of the

**PPC TESTING, INSPECTION AND CERTIFICATION SINGLE
MEMBER S.A.**

(PPC INSPECTRA)

Parameters/ Calibration Item	Range of measurment	Expanded measurement uncertainty (k=2)*	Remarks
Electrical Measurements			
DC Voltage (Measurement)/ Voltage standards, Voltage Calibrators and Sources		parts per 10⁶ output + μV	Calibration according to (i) lab internal procedures, (ii) with standard multimeter. (iii) The calibration may also take place on site in the measurement range [1 – 10kV]
	[1 μV to 100 μV]	0 + 0,029	
	(100 μV to 1 mV]	0 + 0,042	
	(1 mV to 10 mV]	0 + 0,25	
	(10 mV to 120 mV]	5,0 + 0,3	
	(120 mV to 1,2 V]	4,0 + 0,8	
	1,018 V	0 + 0,74	
	(1,2 V to 12 V]	4,0 + 0,5	
	10 V	0 + 0,38	
	(12 V to 120 V]	6,0 + 30	
	(120 V to 1050 V]	6,0 + 100	
DC Voltage (Generation)/ Voltage Measuring Equipment		parts per 10⁶ output + μV	Calibration according to (i) Euramet cg-15 v.3.0:2015,(ii) with standard calibrator
	[1 μV to 100 μV]	0 + 0,01	
	(100 μV to 1 mV]	0 + 0,04	
	(1 mV to 5 mV]	0 + 0,17	
	(5 mV to 10 mV]	0 + 0,34	
	(10 mV to 220 mV]	6,5 + 0,8	
	(220 mV to 2,2 V]	4,0 + 0,8	
	1,018 V	0 + 0,74	
	(2,2 V to 12 V]	3,0 + 2,5	
	10 V	0 + 0,38	

	(12 V to 22 V)	3,0 + 4	
	(22 V to 220 V)	4,0 + 40	
	(220 V to 1100 V)	6,0 + 400	

Parameters/ Calibration Item	Range of measurment	Expanded measurement uncertainty (k=2)*	Remarks
AC Voltage (Measurement)/ Voltage standards, Voltage Calibrators and Sources		parts per 10⁶ output + μV	Calibration according to (i) lab internal procedures, (ii)-with standard multimeter
	(12 mV to 12 V) [1 Hz to 40 Hz]	0,007 + 0,004	
	(40 Hz to 1 kHz]	0,007 + 0,002	
	(1 kHz to 20 kHz]	0,014 + 0,002	
	(20 kHz to 50 kHz]	0,03 + 0,002	
	(50 kHz to 100 kHz]	0,08 + 0,002	
	(100 kHz to 300 kHz]	0,3 + 0,01	
	(300 kHz to 1 MHz]	1 + 0,01	
	(1 MHz to 2 MHz]	1,5 + 0,01	
	(2 MHz to 4 MHz]	4 + 0,07	
	(4 MHz to 8 MHz]	4 + 0,08	
	(8 MHz to 10 MHz]	15 + 0,1	
	(12 V to 120 V) [1 Hz to 40 Hz]	0,02 + 0,004	Calibration (i) according to lab internal procedures, (ii)-with standard multimeter and standard AC divider. (iii) The calibration may also take place on site
	(40 Hz to 1 kHz]	0,02 + 0,002	
	(1 kHz to 20 kHz]	0,02 + 0,002	
	(20 kHz to 50 kHz]	0,035 + 0,002	
	(50 kHz to 100 kHz]	0,12 + 0,002	
	(100 kHz to 300 kHz]	0,4 + 0,01	
	(300 kHz to 1 MHz]	1,5 + 0,01	
	(120 V to 700 V) [1 Hz to 40 Hz]	0,04 + 0,004	
	(40 Hz to 1 kHz]	0,04 + 0,002	
	(1 kHz to 20 kHz]	0,06 + 0,002	
	(20 kHz to 50 kHz]	0,12 + 0,002	
	(50 kHz to 100 kHz]	0,3 + 0,002	
		% Reading	
	[0,7 kV to 3 kV) [50 Hz to 60 Hz]	0,58	
	[3 kV to 7 kV) [50 Hz to 60 Hz]	0,21	
		parts per 10⁶ output + μV	
	[220 μV to 2,2 mV) [10 Hz to 20 Hz]	230 + 4	

Parameters/ Calibration Item	Range of measurment	Expanded measurement uncertainty (k=2)*	Remarks
AC Voltage (Generation)/ Voltage Measuring Equipment	(20 Hz to 40 Hz]	87 + 4	Calibration according to (i) Euramet cg-15 v.3.0:2015, (ii) with standard calibrator
	(40 Hz to 20 kHz]	77 + 4	
	(20 kHz to 50 kHz]	190 + 4	
	(50 kHz to 100 kHz]	480 + 5	
	(100 kHz to 300 kHz]	1000 + 10	
	(300 kHz to 500 kHz]	1300 + 20	
	(500 kHz to 1 MHz]	2600 + 20	
	[22 mV to 220 mV] [10 Hz to 20 Hz]	230 + 12	
	(20 Hz to 40 Hz]	87 + 7	
	(40 Hz to 20 kHz]	77+ 7	
	(20 kHz to 50 kHz]	190 + 7	
	(50 kHz to 100 kHz]	440 + 17	
	(100 kHz to 300 kHz]	800 + 20	
	(300 kHz to 500 kHz]	1300 + 25	
	(500 kHz to 1 MHz]	2600 + 45	
	(220 mV to 2,2 V] [10 Hz to 20 Hz]	230 + 40	
	(20 Hz to 40 Hz]	85 + 15	
	(40 Hz to 20 kHz]	42 + 8	
	(20 kHz to 50 kHz]	73 + 10	
	(50 kHz to 100 kHz]	107 + 30	
	(100 kHz to 300 kHz]	380 + 80	
	(300 kHz to 500 kHz]	950 + 200	
	(500 kHz to 1 MHz]	1600 + 300	

Parameters/ Calibration Item	Range of measurment	Expanded measurement uncertainty (k=2)*	Remarks
		parts per 10⁶ output + μV	
	(2,2 V to 22 V) [10 Hz to 20 Hz]	230 + 400	
	(20 Hz to 40 Hz)	85 + 150	
	(40 Hz to 20 kHz)	42 + 50	
	(20 kHz to 50 kHz)	73 + 100	
	(50 kHz to 100 kHz)	97 + 200	
	(100 kHz to 300 kHz)	270 + 600	
	(300 kHz to 500 kHz)	900 + 2000	
	(500 kHz to 1 MHz)	1400 + 3200	
(continued) AC Voltage (Generation)/ Voltage Measuring Equipment	(22 V to 220 V) [10 Hz to 20 Hz]	230 + 4000	Calibration according to (i) Euramet cg-15 v.3.0:2015, (ii) with standard calibrator
	(20 Hz to 40 Hz)	85 + 1500	
	(40 Hz to 20 kHz)	50 + 600	
	(20 kHz to 50 kHz)	77 + 1000	
	(50 kHz to 100 kHz)	140 + 2500	
	(100 kHz to 300 kHz)	850 + 16000	
	(300 kHz to 500 kHz)	4300 + 40000	
	(500 kHz to 1 MHz)	7500 + 80000	
	(220 V to 1100 V) [50 Hz to 1 kHz]	65 + 3500	
	(1 kHz to 20 kHz)	135 + 6000	
	(20 kHz to 30 kHz)	440 + 11000	
	to 750 V] (30 kHz to 50 kHz)	440 + 11000	
	(50 kHz to 100 kHz)	1600 + 45000	
DC Current (Measurement)/ Current Calibrator and Sources		parts per 10⁶ Reading + parts per 10⁶ Range	Calibration according to (i) lab internal procedures, (ii) with standard multimeter and standard Shunt resistances
	[10 nA to 120 nA)	30 + 400	
	[0,12 μ A to 1,2 μ A)	20 + 40	
	[1,2 μ A to 12 μ A)	20 + 10	
	[12 μ A to 120 μ A)	20 + 8	
	[0,12 mA to 1,2 mA)	20 + 5	
	[1,2 mA to 12 mA)	20 + 5	
	[12 mA to 120 mA)	35 + 5	
	[0,12 A to 1,05 A)	110 + 10	
	[1,05 A to 200 A]	240 + 0	
DC Current (Generation)/ Current measuring Equipment		parts per 10⁶ output + nA	Calibration according to (i) Euramet cg-15 v.3.0:2015, (ii) with standard calibrator and standard current coil multiplier
	[10 μ A to 220 μ A)	37 + 6	
	[0,22 mA to 2,2 mA)	33 + 7	
	[2,2 mA to 22 mA)	33 + 40	
	[22 mA to 220 mA)	42 + 700	
	[0,22 A to 2,2 A)	70 + 12000	
	[2,2 A to 11 A]	350 + 480000	

Parameters/ Calibration Item	Range of measurment	Expanded measurement uncertainty (k=2)*	Remarks
Current Clamp Type Measuring Equipment		% Reading	
	[10 A to 1000 A]	0,33	
AC Current (Measurement)/ Current Calibrator and Sources		% Reading + % Range	Calibration according to (i) lab internal procedure, (ii) with standard multimeter and standard Shunt/ Rogowski. (iii) The calibration may also take place on site
	[10 μ A to 120 μ A]	[10 Hz to 20 Hz) [20 Hz to 45 Hz) [45 Hz to 1 kHz]	
		0,4 + 0,03 0,15 + 0,03 0,06 + 0,03	
	[0,12 mA to 120 mA]	[10 Hz to 20 Hz) [20 Hz to 45 Hz) [45 Hz to 100 Hz) [100 Hz to 5 kHz) [5 kHz to 20 kHz) [20 kHz to 50 kHz) [50 kHz to 100 kHz]	
		0,4 + 0,02 0,15 + 0,02 0,06 + 0,02 0,03 + 0,02 0,06 + 0,02 0,4 + 0,04 0,55 + 0,15	
	[120 mA to 1,05 A]	[10 Hz to 20 Hz) [20 Hz to 45 Hz) [45 Hz to 100 Hz) [100 Hz to 5 kHz) [5 kHz to 20 kHz) [20 kHz to 50 kHz]	
		0,4 + 0,02 0,16 + 0,02 0,08 + 0,02 0,1 + 0,02 0,3 + 0,02 1 + 0,04	
	[20 A to 100 A]	50 Hz	
		3,5 + 0	
	[100 A to 700 A]	50 Hz	
		0,76 + 0	
	[700 A to 1000 A]	50 Hz	
		1,89 + 0	
	[1 kA to 10 kA]	[50 to 60 Hz]	
		1,10 + 0	
AC Current (Generation)/ Current measuring Equipment		parts per 10⁶ output + nA	Calibration according to (i) Euramet cg-15 v.3.0:2015, (ii) with standard calibrator and standard current coil multiplier
	[0,22 mA to 2,2 mA]	[10 Hz to 20 Hz) [20 Hz to 40 Hz) [40 Hz to 1 kHz) [1 kHz to 5 kHz) [5 kHz to 10 kHz]	
		240 + 16 150 + 10 115 + 8 270 + 12 1000 + 65	
	[2,2 mA to 22 mA]	[10 Hz to 20 Hz) [20 Hz to 40 Hz) [40 Hz to 1 kHz) [1 kHz to 5 kHz) [5 kHz to 10 kHz]	
		240 + 40 150 + 35 115 + 35 190 + 110 1000 + 650	

	[0,22 mA to 2,2 mA) [10 Hz to 20 Hz) [20 Hz to 40 Hz) [40 Hz to 1 kHz) [1 kHz to 5 kHz) [5 kHz to 10 kHz]	240 + 400 150 + 350 115 + 350 190 + 550 1000 + 5000	
Parameters/ Calibration Item	Range of measurment	Expanded measurement uncertainty (k=2)*	Remarks
(continued) AC Current (Generation)/ Current Measuring Equipment		parts per 10⁶ output + μA	Calibration according to (i) Euramet cg-15 v.3.0:2015, (ii) with standard calibrator and standard current coil multiplier
	[22 mA to 220 mA) [10 Hz to 20 Hz) [20 Hz to 40 Hz) [40 Hz to 1 kHz) [1 kHz to 5 kHz) [5 kHz to 10 kHz]	240 + 4 150 + 3,5 115 + 2,5 190 + 3,5 1000 + 10	
	[220 mA to 2,2 A) [20 Hz to 1 kHz) [1 kHz to 5 kHz) [5 kHz to 10 kHz]	250 + 35 420 + 80 6500 + 160	
	[2,2 A to 11 A) [40 Hz to 1 kHz) [1 kHz to 5 kHz) [5 kHz to 10 kHz]	440 + 170 900 + 380 3500 + 750	
Current Clamp Type Measuring Equipment		% Reading	
	[10 A to 100 A) 50 Hz (50 Hz to 400 Hz]	3,2 8,4	
	[100 A to 400 A) 50 Hz (50 Hz to 400 Hz]	0,32 0,86	
	(400 A to 1000 A) 50 Hz	0,08	
Resistance DC (Generation)/ Resistance Measuring Equipment		parts per 10⁶	Calibration according to(i) Euramet cg-15 v.3.0:2015, (ii) with standard shunt resistances or standard calibrator
	0 Ω	40 $\mu\Omega$	
	[1 $\mu\Omega$ to 10 $\mu\Omega$]	42	
	(10 $\mu\Omega$ to 100 $\mu\Omega$]	23	
	(100 $\mu\Omega$ to 1 m Ω]	19	
	(1 m Ω to 100 m Ω]	13	
	(100 m Ω to 1,9 Ω]	85	
	(1,9 Ω to 19 Ω]	22	
	(19 Ω to 190 Ω]	9,5	
	(190 Ω to 1,9 k Ω]	8,0	
	(1,9 k Ω to 10 k Ω]	85	
	(10 k Ω to 19 k Ω]	8,0	
	(19 k Ω to 190 k Ω]	10	
	(190 k Ω to 1 M Ω]	17	
	(1 M Ω to 1,9 M Ω]	18	
	(1,9 M Ω to 10 M Ω]	34	
	(10 M Ω to 19 M Ω]	42	
	(19 M Ω to 100 M Ω]	100	
	1 G Ω	5900	
	10 G Ω	1100	

	100 GΩ			4800	
Parameters/ Calibration Item	Range of measurement			Expanded measurement uncertainty (k=2)*	Remarks
Resistance DC (Measurement) / Standards Resistors				parts per 10⁶	
	[1 μΩ to 10 μΩ]			3000	Calibration according to (i) lab internal procedure (ii) with standard calibrator, standard multimeter and shunt resistances
	[10 μΩ to 100 μΩ]			380	
	[100 μΩ to 1 mΩ]			180	
	[1 mΩ to 10 mΩ]			60	
	[10 mΩ to 100 mΩ]			10	
	[100 mΩ to 1 Ω]			7,3	
	(1 Ω to 10 Ω]			7,3	
	(10 Ω to 100 Ω)			11	
	[100 Ω to 1 kΩ]			11	
	[1 kΩ to 10 kΩ]			11	
	(10 kΩ to 100 kΩ]			11	
	(100 kΩ to 1 MΩ]			10	
	(1 MΩ to 10 MΩ]			38	
	(10 MΩ to 100 MΩ]			79	
	(100 MΩ to 1 GΩ]			5900	
	(1 GΩ to 10 GΩ]			500	Calibration according to (i) lab internal procedure, (ii) with standard calibrator and standard multimeter
	(10 GΩ to 100 GΩ]			4660	
Resistance AC (Measurement) / Standards Resistors	10 Ω	[220 μA to 2,2 mA]	[50 Hz to 125 Hz]	290	Calibration according to (i) lab internal procedures, (ii) with standard calibrator and standard multimeter
	[100 Ω to 10 kΩ]			180	
	1 Ω	[2,2 mA to 22 mA]		290	
	[10 Ω to 1 kΩ]			180	
	100 mΩ	[22 mA to 220 mA]		240	
	[1 Ω to 100 Ω]			180	
	10 mΩ	[220 mA to 2,2 A]		340	
	[100 mΩ to 10 Ω]			300	
Resistance AC (Generation)/ Resistance Measuring Equipment	[10 mΩ to 100 mΩ]	[2,2 A to 11 A]	[50 Hz to 125 Hz]	470	Calibration according to (i) lab internal procedures, (ii) with standard resistances
	10 mΩ	10 A		1600	
	100 mΩ	3 A		840	
	1 Ω	1,4 A		12000	
	10 Ω	100 mA		1300	
	100 Ω	20 mA		680	
	1 kΩ	10 mA		680	
	10 kΩ	3 mA		790	

* Where the expanded uncertainty (with 95% coverage) is accompanied by the corresponding unit, it is absolute, while where it is not accompanied by a unit, it is relative.

The Calibration Measurement Capability (CMC), includes the measured quantity, the measurement range and the measurement uncertainty, expressing the minimum measurement uncertainty which can be achieved in a calibration.

Permanent laboratory premises address: **9 Leontariou Str, 15351 Kantza, Pallini, Attiki, Greece.**

Approved signatories: **M. Valsamakis, A. Petrakos, K. Stampolas.**

This Scope of Accreditation replaces the previous one dated 24.01.2025.

The Accreditation Certificate No. **90-9** to ELOT EN ISO/IEC 17025: 2017, is valid until 02.07.2027.

Athens, 23.06.2025

